

Exploring the Principles of GC-MS: Techniques and Applications

Hassan H. Khalifea*, Noor M. Ali

Nahrain University, Iraq

DOI:

<https://doi.org/10.47134/pslse.v2i3.388>

*Correspondence: Hassan H. Khalifea

Email: hasn.hh570@gmail.com

Received: 09-02-2025

Accepted: 09-03-2025

Published: 08-04-2025



Copyright: © 2025 by the authors. Submitted for open access publication under the terms and conditions of the Creative Commons Attribution (CC BY) license

(<http://creativecommons.org/licenses/by/4.0/>).

Abstract: GC-MS operates as a robust analytical method that chemistry labs commonly use because it effectively separates chemicals while identifying their components with precise accuracy. Research instruments like GC-MS provide high quality separation and molecular identification which makes them indispensable for environmental studies as well as pharmaceutical insights and forensic investigations and quality control applications. Extensive usage of GC-MS remains incomplete because researchers must develop an organized framework which illustrates system elements and their supporting functions together with application boundaries. The main objective of this research is to establish a systematic examination of GC-MS core functioning and equipment alongside examining operational details and system constraints. The research shows how GC-MS components including injectors, columns, ionization techniques, mass analyzers and detectors function together to achieve optimal performance results. The research presents current uses of GC-MS in environmental monitoring and pharmaceutical analysis and forensic science through methodological examples while supplying reference spectra. This comprehensive study combines GC-MS principles with practice to provide a unified framework which includes reviews regarding recently developed hyphenated methods and ionization approaches and column development. The research results function as a useful guide for scientists as well as practitioners to help them select systems properly and develop methods and advance GC-MS systems for intricate analytical tasks.

Keywords: GC-MS, Techniques, Applications

Introduction

Despite the numerous analytical techniques available, gas chromatography-mass spectrometry (GC-MS) remains popular, highly reliable, and widely used throughout food, environmental, forensic, and chemical industries (Wang et al, 2020). There are two ways to gain useful information from the analysis of the components eluting from a gas chromatograph (Valdez, 2022).

First, the components may be isolated, and following suitable work-up procedures, their structures can be determined. After the structure is known, the identification of unknowns within that same material must be based on the structural fragment ion pattern

from that same component, unless the complete or partial library spectra can provide the correct identification (Teonata et al, 2021).

Second, the components from a gas chromatograph can be transferred into a mass spectrometer wherein they can impart their mass spectral characteristics (Shi et al., 2020). By analyzing these characteristics, researchers can discern valuable information about the components, such as their molecular weight, fragmentation patterns, and even their isotopic ratios, which can be crucial in determining the compound's origin or verifying its authenticity. With the advancements in technology, GC-MS systems have become increasingly powerful and versatile, allowing for improved sensitivity, resolution, and the ability to analyze a wide range of compounds. Additionally, the advent of hyphenated techniques, such as GC-MS/MS (tandem mass spectrometry) and GCxGC-MS (comprehensive two-dimensional gas chromatography-mass spectrometry), has further expanded the capabilities of GC-MS, enabling the detection and identification of even trace amounts of compounds in complex samples (Russ-Eft et al, 2024). These advancements have solidified GC-MS as a cornerstone technique in analytical chemistry and have made it an indispensable tool in various fields, including environmental monitoring, food safety, forensic investigations, and quality control in the chemical industry (Ranjan et al., 2023). The future of GC-MS looks promising, with ongoing research efforts focused on enhancing instrument performance, developing novel sample preparation techniques, and exploring new applications in emerging fields. As technology continues to evolve, GC-MS will undoubtedly play a vital role in advancing our understanding of the chemical composition of complex matrices and addressing the analytical challenges of the modern world (Putri et al, 2022).

Gas chromatography is a fundamental technique of GC-MS. This involves the separation of components of a sample into individual peaks with a defined retention time. The separation relies on the volatility of the compounds and their interaction with the stationary phase within the column (Meziani et al, 2022). The column is regulated by the flow rate of the carrier gas and a programmed temperature regime to facilitate the separation of compounds. The flow rate is the velocity at which the carrier gas passes from the column inlet to where the column ends in the detector, mainly the mass spectrometer interface. The temperature program is a series of time and temperature instants, called temperature ramps, in which conditions change to allow the volatilization and the use of a differential volatility of the compounds within the mixture. The importance of optimal separation is paramount for the identification of by-products, impurities, and degradation products in excipients and active pharmaceutical ingredients. The combination of GC-MS using full scan functionality and library search software has been shown to be effective and efficient at identifying compounds (Martínez-Huitle et al, 2023). Efficiently packed columns can have theoretical plate counts in the range of 75,000 and high-quality capillary columns up to 2 million. This efficiency can be expressed by two key variables: N , the total number of plates, and H , the height equivalent to a theoretical plate (Marriott et al, 2021). H is calculated by dividing the column length by the total number of plates or dividing the

column by the number of theoretical plates. N increases with an increase in capillary column length, but the number of plates per column is important as well. Key separating factors for a GC column are that it provides efficient separation and good peak shape and reproducibility. These factors are described in more detail in the section below, which covers resolutions, efficiencies, and selectivity (Lago-Rivera et al, 2021).

Methodology

The study methodology uses literature-based studies and technical analysis of Gas Chromatography-Mass Spectrometry (GC-MS) to examine fundamental principles and instrumentation and practical application aspects. Synthesis of detailed information occurred through a qualitative descriptive methodology from peer-reviewed academic sources and scientific manuals and technical documents which explain GC-MS operational mechanisms and performance parameters. The research analyzed secondary data sources about gas chromatography and mass spectrometry device structures which incorporated sample injectors, chromatographic columns, ionization approaches, mass analyzers alongside detectors. Scientists evaluated both separation techniques and detection procedures in GC-MS platforms and their hardware configuration approaches that improve analysis accuracy while maintaining reproducibility standards. Scientific publications and technical standards were extensively used to gain precise up-to-date information about GC-MS systems. The application review section included environmental analysis techniques and pharmaceutical quality control as well as forensic science to showcase the real-world operational value of this technology. This paper evaluates critical GC-MS system factors including selection methods for columns and stationary phases along with carrier gas optimization and detector sensitivity to explain system configurations for multiple research and industrial applications. All interpretations within the study were researched using validated data sources that formed the solid base of discussion and implications. The document-based methodological framework authorizes a complete assessment of GC-MS in its role as an analytical instrument with multiple capabilities.

Result and Discussion

To the point at which the component has zero concentration. As the component diffuses into the stationary phase, it encounters the different internal surfaces of the column material, which leads to selective interactions. The different components have varying affinities for the stationary phase, resulting in differential adsorption. This causes the components to interact with the stationary phase to varying extents and exhibit different retention times. The retention time is the time it takes for a component to travel from the injection port to the detector. The interaction between the components and the stationary phase allows for their separation based on their chemical properties (Kumarajith, 2023). In gas chromatography, the type of stationary phase used plays a crucial role in the separation process. The stationary phase can be a solid support material

or a liquid coated on a solid support. The choice of stationary phase depends on the analytes of interest and the separation requirements. Solid support materials can include porous polymers or powders, while liquid phases can be various types of liquid coatings. The selection of the appropriate stationary phase is crucial as it directly influences the separation efficiency and selectivity of the analytes (Keum et al, 2020). The carrier gas used in gas chromatography also contributes to the separation process. Typically, inert gases such as helium, nitrogen, or hydrogen are used as carrier gases. The choice of carrier gas depends on factors such as the analytes being analyzed, the column dimensions, and the desired analysis speed. The carrier gas carries the sample components through the column and towards the detector. It is important to optimize the carrier gas flow rate to ensure efficient separation and detection (Ghazi et al, 2022). The detector in gas chromatography plays a vital role in analyzing the separated components. Various types of detectors can be used, such as flame ionization detectors (FID), thermal conductivity detectors (TCD), electron capture detectors (ECD), or mass spectrometry detectors (MS). Each detector has its own advantages and limitations, and the choice of detector depends on the analytes being analyzed and the sensitivity required. The detector generates an electrical signal in response to the components eluting from the column, allowing for their identification and quantification (Blumberg, 2021). Gas chromatography is a widely used analytical technique in various fields such as pharmaceuticals, environmental analysis, forensics, and food analysis, among others (Yuan et al, 2022). Its ability to separate and analyze complex mixtures makes it invaluable in research and quality control. The advancements in column technology, stationary phases, carrier gases, and detectors continue to enhance the capabilities of gas chromatography, leading to more accurate and sensitive analyses. In conclusion, gas chromatography is a powerful analytical technique that utilizes the different velocities and interactions between sample components and a stationary phase to achieve separation. Its versatility and reliability make it an essential tool in various scientific disciplines. The continuous advancements in gas chromatography technology contribute to its continued relevance and application in solving complex analytical challenges (Xie et al, 2020).

Column Selection and Stationary Phases

GC columns and stationary phase media are crucial constituents of the configuration of a gas chromatography system. They are the central narrows of separation by the chromatographer. Columns are usually divided into packed and capillary columns; the section discussing the column packing materials concept is included in the next section. Capillary columns, otherwise known as open tubular or empty columns, are the most commonly used, and the introduction of the capillary column changed the course of GC from a technique that could not resolve different classes of volatile compounds to a very powerful and essential separation technique. The application of packed columns is endless, stretching from analyzing ppm levels of pollutants to biogas and fracking (Wei et al, 2023). Stationary phases or adsorbents are the luminal fillers causing the separation of substances,

and their properties are discussed along with the properties and applications of the columns. The dimensions of the columns, like length and I.D., as well as the film thickness of the capillary column, are very important for the speed of analysis, thermal stability, and the quality of the output. Proper polarity, in addition to the required surface activity, is essential for the stationary phase to have a higher binding power for the solute, contributing to a good record-out (Umapathi et al, 2021). The chromatographic factors, like selectivity, efficiency, resolution, solute compound distribution between media, and retention time, as well as the prevailing application to the enduser, are contributing factors in choosing the packing material of the column to be used in analysis. Column-related problems can be appalling, as the cost implication of mismatching the different column units will cascade to the GC-MS module. It is necessary to have adequate information and the correct matching of these column units at the first instance in order to achieve meaningful data and maximum reproducibility for all recurring analyses in GC-MS (Tsuchida et al, 2020).

Basics of Mass Spectrometry Mass spectrometry is employed within GC-MS for analysis, making it essential to understand. The basic concepts involve generating ions and using electric and magnetic fields to manipulate ionized samples before they are detected. These sophisticated techniques are categorized based on the ionization technique used, all applicable in various research contexts (Srivastava et al, 2021). As the mass spectrometry unit suggests, these instruments record and characterize the mass of various ions produced through the ionization techniques. These masses are usually characterized by their corresponding mass-to-charge (m/z) ratios to allow for easy characterization. Most mass spectrometers contain an ion source, a mass analyzer, and a detector. The ion samples are produced and then moved into the mass analyzer, where they are characterized based on their mass. The detector then records the mass of these samples, with the end product being a spectrum showing mass versus intensity, indicating the abundance of these masses. Spectra generated and calibrated from these instruments can be used to elucidate an object's molecular and elemental composition. Data interpretation and calibration procedures are just a couple of processes in mass spectrometry. An understanding of mass analyzers and other ionization techniques discussed in the following sections is imperative for better understanding these processes (Roth et al, 2022).

Ionization Techniques

Ionization is one of the most crucial steps in mass spectrometry. It involves the conversion of a sample to ions so that they can be manipulated and analyzed in the MS system. The choice of ionization technique can greatly influence qualitative and quantitative data. Early mass spectrometers employed electron ionization for the ionization process. Electron ionization is a harsh ionization source that is now being replaced by chemical ionization, fast atom bombardment, and thermospray (Prabhu et al, 2023). There have been some phenomena reported that could affect the ionization efficiency in electron ionization mass spectrometry. For example, cooling the nebulized

droplets in an electron ionization mass spectrometry interface caused a drop in the abundance of detected ions. The reason for this decrease was proposed to be the evaporation of ion-precursor analogs from the droplets that were produced by the ionization of squalene present in the droplets (Liang et al, 2021). Electrospray ionization is widely utilized in existing liquid chromatography-mass spectrometry and its applications. Electrospray ionization ionizes the sample by the application of high voltage. For electrospray ionization to be effective, the sample must be in solution and the solvent must be volatile. Electrospray ionization has been demonstrated to be very effective for the ionization of proteins and peptides, as well as small organic compounds (Li et al, 2021). Although electrospray ionization has extensive advantages, it also has significant limitations. For instance, electrospray ionization is prone to ion suppression when contaminants are sprayed into the mass spectrometry system. Matrix-assisted laser desorption ionization is extensively used for the ionization of biological molecules such as nucleic acids, peptides, and proteins. Currently, besides electrospray ionization and matrix-assisted laser desorption ionization, electrospray ionization-declarative mass spectrometry and atmospheric pressure chemical ionization are being widely used. These four ionization techniques have expanded the use of mass spectrometry in defining many functions of biological systems, such as those in vitamins and alloys. Plasma, as well as some features of human serum. Most research articles and reviews related to mass spectrometry published over the last 10 years utilized one of these four methods. The mechanism and characteristics for both electrospray ionization and matrix-assisted laser desorption ionization will be described in the subsequent section. "Ionization of large molecules" will be described in Part IV. In addition, the applications of various ionization sources in conjunction with various detection devices will be discussed in the next section (Kabir & Furton, 2021).

Gas Chromatograph Components

This subsection emphasizes the elements of the gas chromatograph itself and highlights their features.

As described in Figure 1, It also covers the injector, column, and detector in detail, showing their impact on the separation mechanism. A gas chromatograph consists of at least a means of sample introduction (injector), a capillary column, and a detector plus data system.




Figure 1. A schematic representation of GC-MS instrument

The Injector

The quality of separation using a GC is often dictated by the performance of the chromatograph injector. This is a convenient location at which the sample may be introduced to the chromatograph. A sample solution is typically injected into the flowing carrier gas in the GC inlet, vaporized, and driven onto the column by the carrier gas (Gao et al., 2023). Overfilling the inlet liner by injecting too much sample can lead to sample losses and sample discrimination if the sample discriminates between the liquid phase in the liner and the loading stage onto the column. There are various ways to inject a sample, the simplest being a syringe capable of delivering a known volume of sample apportioned to either manual or automatic injection where the valve controller triggers the injection of the sample through the septum. Split injectors are used when large sample volumes are injected to prevent column contamination with the sample. However, the problem with this injector is venting the excess carrier gas and samples (Chiu & Kuo, 2020).

The Column

The latter is the heart of the gas chromatograph. Its function is to separate the components of the sample mixture, typically into bands of similar compounds, and the actual degree of separation is a quantitative measure of the quality of the chromatogram. The column is enclosed in an oven to maintain the column at an optimal separation temperature, which is another critical method parameter, as will be seen in the next section. Helium and nitrogen are commonly used carrier gases, the former due to abundant sources which mean the gas is cheap to purchase. Nitrogen, while less abundant and less cheap, analysis has shown can also be used as an alternative. Column flow rate must also permit sample loading onto the column while at the same time avoid excessive retention times for the long column used to optimize chromatographic separation (Alseekh et al, 2021).

The Detector

The resolution of the gas chromatography system is determined by its performance and, in particular, the performance of the detector. There are numerous types of detectors available, each with its own unique sensitivities and suitability depending on the sample, analyte, and application (Aliyari & Konermann, 2022). Ultimately, the detector should be in common use for trace level work if at all possible. Variability in the detector performance contributes to the uncertainty in the concentration measurement. Manufacturers supply calibration standards and their own methods, thereby supporting the need for a QC comparison which can demonstrate that the instrumentation is operating within accepted reference criteria. An increase in the number of applications utilizing GC-MS in the food testing sectors will require greater control on the GC-MS detection step with the deployment of hyphenated detectors such as a flame photometric detector or nitrogen phosphorus detector and mass spectrometer, as well as an increase in the number of under lab or field adopted external scientific reports or guidance documents. It is important that laboratories are aware that there are no restrictions on the use of any hyphenated detectors in conjunction with high throughput MS concentration assays.

Conclusion

GC-MS has profoundly influenced multiple domains, including academia, industry, and research, due to its efficiency, portability, and automation, producing swift and reproducible outcomes. Its diverse uses encompass quality control and planetary research, utilizing attributes such as sensitivity, selectivity, and comprehensive separation. GC-MS is essential in medicinal chemistry, pharmaceutical analysis, pharmacognosy, process control, and pharmaceutical biotechnology, demonstrating its significance in chemical synthesis, stability assessment, and impurity characterization. It facilitates research, quality assurance, and the production of diverse pharmaceutical components within the industry. Although computational technologies aid in data analysis, the determination of chemical structures from mass spectra using GC-MS still depends on human experience because of the complexities involved in data interpretation. The continued study of GC-MS capabilities has possibilities for future breakthroughs, although the analysis of complex materials necessitates continuous technological and instrumental growth.

References

- Aliyari, E., & Konermann, L. (2022). Formation of Gaseous Peptide Ions from Electrospray Droplets: Competition between the Ion Evaporation Mechanism and Charged Residue Mechanism. *Analytical Chemistry*.
- Alseekh, S., Aharoni, A., Brotman, Y., & Contrepois, K. (2021). Mass spectrometry-based metabolomics: A guide for annotation, quantification and best reporting practices. *Nature*.
- Blumberg, L. M. (2021). Theory of gas chromatography. *Gas Chromatography*.

- Chiu, H. H., & Kuo, C. H. (2020). Gas chromatography-mass spectrometry-based analytical strategies for fatty acid analysis in biological samples. *Journal of Food and Drug Analysis*.
- Gao, F., Liu, C., Zhang, L., Liu, T., Wang, Z., & Song, Z. (2023). Wearable and flexible electrochemical sensors for sweat analysis: A review. *Microsystems & ...*
- Ghazi, M. G. M., Lee, L. C., Sino, H., & Halim, M. I. A. (2022). Review of contemporary chemometric strategies applied on preparing GC–MS data in forensic analysis. *Microchemical Journal*.
- Kabir, A., & Furton, K. G. (2021). Applications of gas chromatography in forensic science. *Gas Chromatography*.
- Keum, K., Kim, J. W., Hong, S. Y., Son, J. G., & Lee, S. S. (2020). Flexible/stretchable supercapacitors with novel functionality for wearable electronics. *Advanced*.
- Kumarajith, T. (2023). *University of Tasmania Open Access Repository Cover sheet*. <https://utas.edu.au>
- Lago-Rivera, D., Grandi, S., Rakonjac, J. V., & Seri, A. (2021). Telecom-heralded entanglement between multimode solid-state quantum memories. *Nature*.
- Li, C., Chu, S., Tan, S., Yin, X., Jiang, Y., & Dai, X. (2021). Towards higher sensitivity of mass spectrometry: A perspective from the mass analyzers. *Frontiers In*.
- Liang, Y., Li, J., Xue, Y., Tan, T., Jiang, Z., & He, Y. (2021). Benzene decomposition by non-thermal plasma: A detailed mechanism study by synchrotron radiation photoionization mass spectrometry and theoretical calculations. *Journal of Hazardous*.
- Marriott, P. J., Chin, S. T., & Nolvachai, Y. (2021). Techniques and application in comprehensive multidimensional gas chromatography–mass spectrometry. *Journal of Chromatography A*.
- Martínez-Huitle, C. A., Rodrigo, M. A., & Sirés, I. (2023). A critical review on latest innovations and future challenges of electrochemical technology for the abatement of organics in water. *Applied Catalysis B*.
- Meziani, A., Verloy, S., Ferroukhi, O., & Roca, S. (2022). Evaluation of gas chromatography columns with radially elongated pillars as second-dimension columns in comprehensive two-dimensional gas chromatography. *Analytical*.
- Prabhu, G. R. D., Williams, E. R., & Wilm, M. (2023). Mass spectrometry using electrospray ionization. *Nature Reviews Methods*.
- Putri, S. P., Ikram, M. M. M., Sato, A., & Dahlan, H. A. (2022). Application of gas chromatography-mass spectrometry-based metabolomics in food science and technology. *Journal of Bioscience*.
- Ranjan, S., Roy, C., & Sinha, S. K. (2023). Gas chromatography–mass spectrometry (GC–MS): A comprehensive review of synergistic combinations and their applications in the past two decades. *Journal of Analytical*.
- Roth, K., Pemula, L., & Zepeda, J. (2022). Towards total recall in industrial anomaly detection. *Proceedings of The ...*

- Russ-Eft, D., Preskill, H., & Jordan, J. B. (2024). *Evaluation in organizations: A systematic approach to enhancing learning, performance, and change*.
- Shi, L., Bucknall, M. P., Young, T. L., Zhang, M., Hu, L., & Bing, J. (2020). Gas chromatography–mass spectrometry analyses of encapsulated stable perovskite solar cells. *Science*.
- Srivastava, N., Singh, A., Kumari, P., & Nishad, J. H. (2021). Advances in extraction technologies: Isolation and purification of bioactive compounds from biological materials. *Natural Bioactive*.
- Teonata, N., Wijaya, V. A., & Vithaloka, V. S. (2021). An introduction to different types of gas chromatography. *Jurnal Sains Dan*.
- Tsuchida, S., Umemura, H., & Nakayama, T. (2020). Current status of matrix-assisted laser desorption/ionization–time-of-flight mass spectrometry (MALDI-TOF MS) in clinical diagnostic microbiology. *Molecules*.
- Umapathi, R., Sonwal, S., Lee, M. J., & Rani, G. M. (2021). Colorimetric based on-site sensing strategies for the rapid detection of pesticides in agricultural foods: New horizons, perspectives, and challenges. *Coordination Chemistry*.
- Valdez, C. A. (2022). Gas chromatography-mass spectrometry analysis of synthetic opioids belonging to the fentanyl class: A review. *Critical Reviews in Analytical Chemistry*.
- Wang, S., Chen, H., & Sun, B. (2020). Recent progress in food flavor analysis using gas chromatography–ion mobility spectrometry (GC–IMS). *Food Chemistry*.
- Wei, G., Dan, M., Zhao, G., & Wang, D. (2023). Recent advances in chromatography-mass spectrometry and electronic nose technology in food flavor analysis and detection. *Food Chemistry*.
- Xie, S. M., Chen, X. X., Zhang, J. H., & Yuan, L. M. (2020). Gas chromatographic separation of enantiomers on novel chiral stationary phases. *TrAC Trends in Analytical*.
- Yuan, C., Jia, W., Yu, Z., Li, Y., & Zi, M. (2022). Are highly stable covalent organic frameworks the key to universal chiral stationary phases for liquid and gas chromatographic separations? *Journal of the American Chemical Society*.